



## DEVICE DETAILS

|                                      |  |
|--------------------------------------|--|
| <b>NAME OF DEVICE</b>                | FITSCAN 365 PULSE OXIMETER   |
| <b>ESTABLISHMENT NAME</b>            | MAXWELL MEDICAL SUPPLIES SDN BHD   |
| <b>ROLE OF ESTABLISHMENT</b>         | AUTHORIZED REPRESENTATIVE  |
| <b>REGISTRATION NO</b>               | GB1845921-81484  |
| <b>BRAND NAME</b>                    | FITSCAN 365  |
| <b>MEDICAL DEVICE CATEGORY</b>       | MD 1302 - Monitoring devices of vital physiological parameters   |
| <b>DEVICE GROUPING TYPE</b>          | SINGLE   |
| <b>DEVICE DESCRIPTION</b>            | FitScan 365 Pulse Oximeter is used to measure arterial oxygen saturation (%SpO <sub>2</sub> ) of hemoglobin and pulse rate, an important indicator of respiratory function, using a pair of sensors placed on a thin part of the patient's fingertip.          |
| <b>DEVICE INTENDED PURPOSE</b>       | FitScan 365 Pulse Oximeter is intended for spot checking of functional oxygen saturation of arterial hemoglobin (SpO <sub>2</sub> ) and pulse rate. This portable device is indicated for use in adult patients in clinical institution and home environments. |
| <b>VALIDITY DATE OF REGISTRATION</b> | 21/12/2021 - 20/12/2026  |

## LIST OF DEVICE



| NO | NAME OF DEVICE             | IDENTIFIER    |
|----|----------------------------|---------------|
| 1  | FITSCAN 365 PULSE OXIMETER | 0785104813816 |