

DEVICE DETAILS		
NAME OF DEVICE	HANDHELD PULSE OXIMETER (PC-66B)	
<b>ESTABLISHMENT NAME</b>	LEPU COMMERCE(MALAYSIA) SDN.BHD	
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE	
REGISTRATION NO	GC4172524-175148	
BRAND NAME	LEPU MEDICAL	
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters	
DEVICE GROUPING TYPE	SYSTEM	
DEVICE DESCRIPTION	The pulse oximeter provides a non-invasive measurement of the arterial oxygen saturation (SpO2) and the heart rate (pulse rate). Oxygen saturation indicates the percentage of haemoglobin in arterial blood that is loaded with oxygen. Therefore, it is an important parameter for assessing respiratory function. ● 22-inch TFT display for SpO2, PI, Plethysmogram, PR ● MoveOxy SpO2 technology with artefact removal and anti- motion performance ● Various SpO probes for Adult, Paediatric, and Neonate ● Inbuilt lithium battery for continuous monitoring ● Up to 384 hours SpO2 data storage for reviewing ● Adjustable alarm limit, pitch tone level, and key tone	
DEVICE INTENDED PURPOSE	This Handheld Pulse Oximeter is intended for measuring and recording the pulse rate and functional oxygen saturation (SpO2). It is applicable for monitoring SpO2 and pulse rate of adult and paediatric patients in clinical institutions and homes.	
VALIDITY DATE OF REGISTRATION	25/06/2024 - 24/06/2029	

LIST OF DEVICE		
NO	NAME OF DEVICE	IDENTIFIER
1	HANDHELD PULSE OXIMETER	PC-66B
2	ACCESSORIES	(not set)
3	SpO2 probe	KS-C01
4	Holder	KRK112
5	Battery	LARQE
6	Charging cable	LXCP12-005
7	Data cable	USB-A