



## DEVICE DETAILS

<b>NAME OF DEVICE</b>	PULSE OXIMETER (PF-10AW,PF-10AW1,PF-10A,PF-10A1, PF-10BW, PF-10BW1?PF-10B,PF-10B1)
<b>ESTABLISHMENT NAME</b>	MURSMEDIC MALAYSIA SDN. BHD.
<b>ROLE OF ESTABLISHMENT</b>	AUTHORIZED REPRESENTATIVE
<b>REGISTRATION NO</b>	GB10766524-191958
<b>BRAND NAME</b>	LEPU MEDICAL
<b>MEDICAL DEVICE CATEGORY</b>	MD 1302 - Monitoring devices of vital physiological parameters
<b>DEVICE GROUPING TYPE</b>	FAMILY
<b>DEVICE DESCRIPTION</b>	<p>The pulse oxygen saturation, the percentage of HbO<sub>2</sub> in the total Hb in the blood, namely the so-called O<sub>2</sub> concentration in the blood, is an important bio-parameter for the respiration. The intergradations of oxygen molecule in the blood and hemoglobin in erythrocyte are reversible. Hemoglobins integrated with oxygen are called HbO<sub>2</sub>, and those released O<sub>2</sub> are called HbR. Oxyhemoglobin and deoxyhemoglobin (HHb) absorb red Marginal and infrared of specific wavelength, thus the Marginal of the two wavelengths emitted from SpO<sub>2</sub> sensor is absorbed when it passes through the finger, and approximate pulse oxygen saturation is calculated by measuring absorption of the Marginal. The device is a lightweight, portable health wrist oximeter for use in the home or in healthcare facilities. SpO<sub>2</sub> measurement technology is based on developed photoelectron method, the circuit design and calculation software was developed by Shenzhen Viatom Technology Co., Ltd. The SpO<sub>2</sub> sensor receives the optical signal from the red light and infra-red light through the finger. Insert the finger into the oximeter, there are two emitting tube (red light diodes and infrared diodes) located on the inner upside of the sensor and they can emit red light and infrared; There is the receiving end located on the inner downside of the sensor, and it can transmit the red light and infrared into the pulse signal through finger. The MCU receives the pulse signal, gets the frequency signal by counting, processes its digital signal, and finally gets the measured SpO<sub>2</sub> value. The PR is averagely calculated by above peak intervals of PR waveform. The device is powered by internal battery. The device is not for life-supporting or lifesustaining, not for implant. The device or sensor is not sterile and the sensor does not need sterilization and the sensor is reusable but does not need re-sterilization since it is not sterile. The device is for prescription. The device does not contain drug or biological products. The device consist of main unit, SpO<sub>2</sub> sensor, wristband and charging cable. The main unit is mainly composed of MCU (built-in Bluetooth module), Power circuit, SpO<sub>2</sub> measurement circuit, Display control circuit, etc. that is an integrated product with modular design</p>
<b>DEVICE INTENDED PURPOSE</b>	This Oximeter is intended for measuring the pulse rate and functional oxygen saturation (SpO <sub>2</sub> ) through a patient's finger. It is applicable for spot-checking SpO <sub>2</sub> and pulse rate of adult and pediatric patients in homes and medical clinics.
<b>VALIDITY DATE OF REGISTRATION</b>	15/12/2024 - 14/12/2029

## LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	Pulse Oximeter	PF?10AW
2	Pulse Oximeter	PF?10AW1
3	Pulse Oximeter	PF?10A
4	Pulse Oximeter	PF?10A1
5	Pulse Oximeter	PF?10BW
6	Pulse Oximeter	PF?10BW1
7	Pulse Oximeter	PF?10B
8	Pulse Oximeter	PF?10B1

