

DEVICE DETAILS		
NAME OF DEVICE	PULSE OXIMETER	
ESTABLISHMENT NAME	REVO HEALTHCARE SDN BHD	
ROLE OF ESTABLISHMENT	MANUFACTURER	
REGISTRATION NO	GC3365525-210246	
BRAND NAME	NOVOPLUS NEUW	
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters	
DEVICE GROUPING TYPE	SINGLE	
DEVICE DESCRIPTION	The NovoPlus Neuw Pulse Oximeter (model: PO77A) is a small, portable, non-invasive, easy-to-use, handheld device that displays oxygen saturation levels and pulse rate. Its specifications include an oxygen saturation measurement range of 0% to 100% and a pulse rate measurement range of 25 bpm to 250 bpm. The PO77A pulse oximeter consists of a probe, electronic circuits, a plastic enclosure, a compartment for battery placement, and a display that shows various information, including the oxygen saturation level, oxygen saturation symbol, low battery indicator, pulse rate, plethysmogram, and more. The device is powered by two AAA alkaline batteries, and it emits a sound when blood oxygen falls below or pulse rate exceeds the set threshold.	
DEVICE INTENDED PURPOSE	The NovoPlus Neuw Pulse Oximeter is a reusable device and is intended for spot-checking of oxygen saturation and pulse rate for use with the finger of adult patients.	
VALIDITY DATE OF REGISTRATION	02/09/2025 - 01/09/2030	

LIST OF DEVICE			
NO	NAME OF DEVICE	IDENTIFIER	
1	PULSE OXIMETER	PO77A	