



DEVICE DETAILS

NAME OF DEVICE	FINGERTIP PULSE OXIMETER
ESTABLISHMENT NAME	QUANTUM UPSTREAM SDN.BHD.
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE
REGISTRATION NO	GC5277326-222891
BRAND NAME	EZYSWISS
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters
DEVICE GROUPING TYPE	SINGLE
DEVICE DESCRIPTION	ezyswiss™ Fingertip Pulse Oximeter main function of the Pulse Oximeter is to measure the functional oxygen saturation of arterial hemoglobin (SpO ₂) and pulse rate as non-invasive spot checking monitoring. The subject device determines the functional oxygen saturation of arterial hemoglobin (SpO ₂) by measuring the absorption of red and infrared light passing through perfused tissue. Changes in absorption caused by the pulsation of blood in the vascular bed are used to determine SpO ₂ reading and pulse rate. The characteristic of the device is a stand alone device with 1 AAA Alkaline battery. It includes an LED, OLED display screen and a warning or indicator function for abnormal readings.
DEVICE INTENDED PURPOSE	Measurement of functional arterial oxygen saturation (SpO ₂) and pulse rate.
VALIDITY DATE OF REGISTRATION	16/01/2026 - 15/01/2031

LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
No results found.		