



## DEVICE DETAILS

<b>NAME OF DEVICE</b>	FINGERTIP PULSE OXIMETER
<b>ESTABLISHMENT NAME</b>	QUANTUM UPSTREAM SDN.BHD.
<b>ROLE OF ESTABLISHMENT</b>	AUTHORIZED REPRESENTATIVE
<b>REGISTRATION NO</b>	GC5277326-222891
<b>BRAND NAME</b>	EZYSWISS
<b>MEDICAL DEVICE CATEGORY</b>	MD 1302 - Monitoring devices of vital physiological parameters
<b>DEVICE GROUPING TYPE</b>	SINGLE
<b>DEVICE DESCRIPTION</b>	ezyswiss™ Fingertip Pulse Oximeter main function of the Pulse Oximeter is to measure the functional oxygen saturation of arterial hemoglobin (SpO <sub>2</sub> ) and pulse rate as non-invasive spot checking monitoring. The subject device determines the functional oxygen saturation of arterial hemoglobin (SpO <sub>2</sub> ) by measuring the absorption of red and infrared light passing through perfused tissue. Changes in absorption caused by the pulsation of blood in the vascular bed are used to determine SpO <sub>2</sub> reading and pulse rate. The characteristic of the device is a stand alone device with 1 AAA Alkaline battery. It includes an LED, OLED display screen and a warning or indicator function for abnormal readings.
<b>DEVICE INTENDED PURPOSE</b>	Measurement of functional arterial oxygen saturation (SpO <sub>2</sub> ) and pulse rate.
<b>VALIDITY DATE OF REGISTRATION</b>	16/01/2026 - 15/01/2031

## LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
No results found.		