

DEVICE DETAILS

NAME OF DEVICE	FINGERTIP PULSE OXIMETER
ESTABLISHMENT NAME	QUANTUM UPSTREAM SDN.BHD.
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE
REGISTRATION NO	GC5277326-222891
BRAND NAME	EZYSWISS
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters
DEVICE GROUPING TYPE	SINGLE
DEVICE DESCRIPTION	ezyswissTM Fingertip Pulse Oximeter main function of the Pulse Oximeter is to measure the functional oxygen saturation of arterial hemoglobin (SpO ₂) and pulse rate as non-invasive spot checking monitoring. The subject device determines the functional oxygen saturation of arterial hemoglobin (SpO ₂) by measuring the absorption of red and infrared light passing through perfused tissue. Changes in absorption caused by the pulsation of blood in the vascular bed are used to determine SpO ₂ reading and pulse rate. The characteristic of the device is a stand alone device with 1 AAA Alkaline battery. It includes an LED, OLED display screen and a warning or indicator function for abnormal readings.
DEVICE INTENDED PURPOSE	Measurement of functional arterial oxygen saturation (SpO ₂) and pulse rate.
VALIDITY DATE OF REGISTRATION	16/01/2026 - 15/01/2031

LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
No results found.		