



## DEVICE DETAILS

<b>NAME OF DEVICE</b>	PULSE OXIMETER
<b>ESTABLISHMENT NAME</b>	AA ALIVE SDN BHD
<b>ROLE OF ESTABLISHMENT</b>	MANUFACTURER
<b>REGISTRATION NO</b>	GB10889026-231055
<b>BRAND NAME</b>	IMDK
<b>MEDICAL DEVICE CATEGORY</b>	MD 1302 - Monitoring devices of vital physiological parameters
<b>DEVICE GROUPING TYPE</b>	SINGLE
<b>DEVICE DESCRIPTION</b>	The device intended for use in measuring and displaying functional oxygen saturation of arterial hemoglobin (%SpO <sub>2</sub> ) and pulse rate (PR). The device measures SpO <sub>2</sub> and PR with a SpO <sub>2</sub> sensor and displays on the display screen after certain further processing, it can be used to measure human Hemoglobin Saturation and heart rate through finger.
<b>DEVICE INTENDED PURPOSE</b>	The Pulse Oximeter is a portable non-invasive device intended for use in measuring and displaying oxygen saturation of arterial hemoglobin (SpO <sub>2</sub> ) and pulse rate of adults through the finger in hospitals, hospital-type facilities, and home environments. The patient is an intended operator
<b>VALIDITY DATE OF REGISTRATION</b>	16/04/2026 - 15/04/2031

## LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
No results found.		