



## DEVICE DETAILS

NAME OF DEVICE	FINGERTIP PULSE OXIMETER
ESTABLISHMENT NAME	GAS WALKER MIRAMED SDN BHD
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE
REGISTRATION NO	GC9480821-57718
BRAND NAME	YUWELL
MEDICAL DEVICE CATEGORY	MD 1100 - GENERAL ACTIVE MEDICAL DEVICES
DEVICE GROUPING TYPE	FAMILY
DEVICE DESCRIPTION	Pulse oximeter is a noninvasive method for monitoring a person's oxygen saturation (SO <sub>2</sub> ). Though its reading of SpO <sub>2</sub> (peripheral oxygen saturation) is not always identical to the more desirable reading of SaO <sub>2</sub> (arterial oxygen saturation) from arterial blood gas analysis, the two are correlated well enough that the safe, convenient, noninvasive, inexpensive pulse oximetry method is valuable for measuring oxygen saturation in clinical use.
DEVICE INTENDED PURPOSE	The fingertip Oximeter can be used to measure human Haemoglobin Saturation and heart rate through finger, the product is suitable for family, hospital (including clinical use in internist/surgery, Anaesthesia, paediatrics, intensive sports (It can be used before or after sport. Operation in sport is not recommended) and etc.The product is not suitable to monitor patient continuously.
VALIDITY DATE OF REGISTRATION	24/03/2021 - 23/03/2026



## LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	Fingertip Pulse Oximeter	YX302
2	Fingertip Pulse Oximeter	YX100