



DEVICE DETAILS

NAME OF DEVICE	CHECKME O2 PULSE OXIMETER
ESTABLISHMENT NAME	BMEC (MALAYSIA) SDN. BHD
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE
REGISTRATION NO	GB3693921-57628
BRAND NAME	OXIBAND
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters
DEVICE GROUPING TYPE	SYSTEM
DEVICE DESCRIPTION	The CheckMe O2 Pulse Oximeter measures the oxygen saturation and pulse rate. It works by shining two beams into the small blood vessels or capillaries of finger, reflecting the amount of oxygen in the blood and displaying the measurement.
DEVICE INTENDED PURPOSE	The Oxiband Pulse Oximeter is a wrist pulse oximeter indicated for use in measuring, displaying, storing and transmitting functional oxygen saturation of arterial hemoglobin (SpO2) and pulse rate for adult patients. The Oxiband it is intended for spot-check and/ or continuous data collection, and not continuous monitoring. It can be used in sleep labs, long-term care, hospitals and home use.
VALIDITY DATE OF REGISTRATION	10/03/2021 - 09/03/2026

LIST OF DEVICE

MEDICAL DEVICE AUTHORITY

NO	NAME OF DEVICE	IDENTIFIER
1	Pulse Oximeter	Oxiband
2	SpO2 Ring Sensor	FP10R
3	USB Charging Cable	540-00240-00