



## DEVICE DETAILS

<b>NAME OF DEVICE</b>	CHECKME O2 PULSE OXIMETER
<b>ESTABLISHMENT NAME</b>	BMEC (MALAYSIA) SDN. BHD
<b>ROLE OF ESTABLISHMENT</b>	AUTHORIZED REPRESENTATIVE
<b>REGISTRATION NO</b>	GB3693921-57628
<b>BRAND NAME</b>	OXIBAND
<b>MEDICAL DEVICE CATEGORY</b>	MD 1302 - Monitoring devices of vital physiological parameters
<b>DEVICE GROUPING TYPE</b>	SYSTEM
<b>DEVICE DESCRIPTION</b>	The CheckMe O2 Pulse Oximeter measures the oxygen saturation and pulse rate. It works by shining two beams into the small blood vessels or capillaries of finger, reflecting the amount of oxygen in the blood and displaying the measurement.
<b>DEVICE INTENDED PURPOSE</b>	The Oxiband Pulse Oximeter is a wrist pulse oximeter indicated for use in measuring, displaying, storing and transmitting functional oxygen saturation of arterial hemoglobin (SpO2) and pulse rate for adult patients. The Oxiband it is intended for spot-check and/ or continuous data collection, and not continuous monitoring. It can be used in sleep labs, long-term care, hospitals and home use.
<b>VALIDITY DATE OF REGISTRATION</b>	10/03/2021 - 09/03/2026

## LIST OF DEVICE

MEDICAL DEVICE AUTHORITY

NO	NAME OF DEVICE	IDENTIFIER
1	Pulse Oximeter	Oxiband
2	SpO2 Ring Sensor	FP10R
3	USB Charging Cable	540-00240-00