



DEVICE DETAILS

NAME OF DEVICE	PULSE OXIMETER
ESTABLISHMENT NAME	ONEWORLD CORPORATE (M) SDN. BHD.
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE
REGISTRATION NO	GC7968521-77190
BRAND NAME	SURGIPLUS/M70C
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters
DEVICE GROUPING TYPE	FAMILY
DEVICE DESCRIPTION	The pulse oximeter provides a non-invasive measurement of the arterial oxygen saturation (SpO ₂) and the heart rate (pulse rate). Oxygen saturation indicates the percentage of haemoglobin in arterial blood that is loaded with oxygen. Oximeters work by the principles of spectrophotometry: the relative absorption of red (absorbed by deoxygenated blood) and infrared (absorbed by oxygenated blood) light of the systolic component of the absorption waveform correlates to arterial blood oxygen saturation.
DEVICE INTENDED PURPOSE	The pulse oximeter provides a non-invasive measurement of the arterial oxygen saturation (SpO ₂) and the heart rate (pulse rate)
VALIDITY DATE OF REGISTRATION	01/11/2021 - 31/10/2026

LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	PULSE OXIMETER	9550201308180
2	PULSE OXIMETER	9550201308180-1
3	PULSE OXIMETER	9550201308180-2