



DEVICE DETAILS

NAME OF DEVICE	Pulse Oximeter
ESTABLISHMENT NAME	ANTAH SRI RADIN SDN BHD
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE
REGISTRATION NO	GC5157821-78346
BRAND NAME	PROMISE
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters
DEVICE GROUPING TYPE	SINGLE
DEVICE DESCRIPTION	PROMISE Pulse Oximeter, based on all digital technology, is intended for non-invasive spot-check measurement of functional oxygen saturation of arterial hemoglobin (SpO ₂). Advanced DSP algorithm can reduce the influence of motion artifact and improve measurement accuracy of low perfusion.
DEVICE INTENDED PURPOSE	PROMISE Pulse Oximeter is intended for non-invasive spot-check measurement of functional oxygen saturation of arterial hemoglobin (SpO ₂). This is suitable for the clinic, oxygen bar, sports health (using it before or after activity, not advised to use during activity), community health care, and aviation etc.
VALIDITY DATE OF REGISTRATION	29/09/2021 - 28/09/2026

LIST OF DEVICE

MEDICAL DEVICE AUTHORITY

NO	NAME OF DEVICE	IDENTIFIER
1	Pulse Oximeter	PRO-F9