

DEVICE DETAILS		
NAME OF DEVICE	A2 PULSE OXIMETER	
ESTABLISHMENT NAME	STARVILION SDN BHD	
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE	
REGISTRATION NO	GB2673821-80060	
BRAND NAME	HEAL FORCE	
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters	
DEVICE GROUPING TYPE	SINGLE	
DEVICE DESCRIPTION	The A2 Pulse Oximeter is intended for measuring the pulse rate and functional oxygen saturation (SpO2) through patient's finger. It is applicable for spot-checking SpO2 and pulse rate of adult and pediatric patients in homes and clinics. This product is not suitable for monitoring the patient's prolonged use.	
DEVICE INTENDED PURPOSE	This Pulse Oximeter is a kind of innovated medical detection device with non-invasive and continuous features for SPO2 and PR detection. It is portable and easy to measure the SPO2 and PR value quickly and precisely. This can be through the finger Pulse Oximeter to measure human blood oxygen saturation and heart rate. This product is suitable for family, clinic, oxygen bar, sports health (use before and after exercise is not recommended for use during exercise), community health and other ranges. This product is not suitable for monitoring the patient's prolonged use.	
VALIDITY DATE OF REGISTRATION	25/11/2021 - 24/11/2026	

LIST OF DEVICE		
NO	NAME OF DEVICE	IDENTIFIER
1	A2 PULSE OXIMETER	A2