



DEVICE DETAILS

NAME OF DEVICE	FINGERTIP OXIMETER
ESTABLISHMENT NAME	UCREST BERHAD
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE
REGISTRATION NO	GC8852522-92820
BRAND NAME	CREATIVE
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters
DEVICE GROUPING TYPE	SINGLE
DEVICE DESCRIPTION	Fingertip Oximeter (PC-60FW) is a non-invasive, non-sterile portable device intended for measuring the pulse rate and functional oxygen saturation (SpO ₂) through a patient's finger. It is applicable for spot checking of SpO ₂ level and pulse rate of adult and paediatric patients at homes and medical clinics settings. It is lightweight, small in size, low power consumption, convenient operation and portability. It only required patient to place one finger into fingertip photoelectric sensor for diagnosis, and the display screen will directly show measured value of functional oxygen saturation and pulse rate.
DEVICE INTENDED PURPOSE	Fingertip Oximeter is a non-invasive, non-sterile portable device intended for measuring the pulse rate and functional oxygen saturation (SpO ₂) through a patient's finger. It is indicated for spot checking of SpO ₂ level and pulse rate of adult and paediatric patients at homes and medical clinics settings.
VALIDITY DATE OF REGISTRATION	13/05/2022 - 12/05/2027

LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	FINGERTIP OXIMETER	11000207