



## DEVICE DETAILS

<b>NAME OF DEVICE</b>	N95 SMP ANTIBACTERIAL MEDICAL FACE MASK
<b>ESTABLISHMENT NAME</b>	DEXINA FLORAPRO RESOURCES (M) SDN. BHD.
<b>ROLE OF ESTABLISHMENT</b>	MANUFACTURER
<b>REGISTRATION NO</b>	GA3139322-94046
<b>BRAND NAME</b>	TRUVIGOR
<b>MEDICAL DEVICE CATEGORY</b>	SINGLE-USE DEVICES
<b>DEVICE GROUPING TYPE</b>	SINGLE
<b>DEVICE DESCRIPTION</b>	TRADE MARK : TRUVIGOR MODEL : N95 SMP ANTIBACTERIAL MEDICAL FACE MASK RATING EN149 CLASSIFICATION =FFP2 SPLASH RESISTANCE TYPE :IIR [ FACE MASK TYPE RESPIRATORY PROTECTIVE DEVICE TO BE USED IN AND/OR OPERATING THEATRE OTHER MEDICAL SETTINGS] TEST REPORT :2021CE1949 , 2022CE0104, 03-14B-02-2022-39-03-1 , 2022CE010-S1 , 03-14B-02-2022-39-04-2 PRODUCT STANDARD: EN14683:2019 PRODUCT : MEDICAL FACE MASK SHAPE: 3D PROTECTIVE 5 PLY FACE MASK AVAILABILITY :SINGLE USE ONLY COUNTRY OF ORIGIN :MALAYSIA (MADE IN MALAYSIA) MANUFACTURER: DEXINA FLORAPRO RESOURCES (M) SDN BHD FEATURE :ANTIBACTERIAL ,ANTIVIRAL ,DEODORIZATION ,SAFE,COMFORTABLE ,STYLISH
<b>DEVICE INTENDED PURPOSE</b>	TO BE WORN BY USER AND ACT AS PROTECTIVE BARRIER FOR NOSE AND MOUTH FROM LARGE PARTICLE DROPLETS SPLASHES , SPRAYS OR SPLATTER THAT MAY CONTAIN VIRUSES AND BACTERIA OR ANY OTHER AIR - BORNE PARTICULATES .THIS FACE MASK USED AS A SURGICAL MASK .
<b>VALIDITY DATE OF REGISTRATION</b>	08/06/2022 - 07/06/2027

## LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	N95 SMP ANTIBACTERIAL MEDICAL FACE MASK	N95 SMP