

DEVICE DETAILS		
DEVICE DETAILS		
NAME OF DEVICE	MEDISANA PULSE OXIMETER PM 100 CONNECT	
ESTABLISHMENT NAME	ZUELLIG PHARMA SDN.BHD	
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE	
REGISTRATION NO	GC9976721-74944	
BRAND NAME	MEDISANA	
MEDICAL DEVICE CATEGORY	MD 1100 - GENERAL ACTIVE MEDICAL DEVICES	
DEVICE GROUPING TYPE	SINGLE	
DEVICE DESCRIPTION	The Pulse Oximeter PM 100 CONNECT is a portable non-invasive device used to measure and record the oxygen saturation of arterial hemoglobin (SpO2) and pulse rate of users at any moment. It is a small and compact device which is applicable in the hospital and home use with simple and convenient operation. However, it is reminded that the Pulse Oximeter PM 100 CONNECT is not suitable for continuous blood oxygen monitoring. The measuring range for the SpO2 is 70% to 99% with 1% resolution while the measuring range for the pulse rate is 30 to 235 beats per minute (bpm) with a resolution of 1 bpm.	
DEVICE INTENDED PURPOSE	The Pulse Oximeter PM 100 CONNECT is intended for spot-checking of oxygen saturation of arterial hemoglobin (SpO2) and pulse rate users. It is not suitable for continuous monitoring.	
VALIDITY DATE OF REGISTRATION	20/09/2021 - 19/09/2026 MALAYSIA	
MED	ICAL DEVICE ALITHORITY	

LIST OF DEVICE		
NO	NAME OF DEVICE	IDENTIFIER
1	MEDISANA PULSE OXIMETER PM 100 CONNECT	48506