



DEVICE DETAILS

NAME OF DEVICE	3D Air Flip Medical Face Mask 4PLY
ESTABLISHMENT NAME	JINGGA ANGGUN SDN BHD
ROLE OF ESTABLISHMENT	MANUFACTURER
REGISTRATION NO	GA2343823-133955
BRAND NAME	MIOS
MEDICAL DEVICE CATEGORY	SINGLE-USE DEVICES
DEVICE GROUPING TYPE	FAMILY
DEVICE DESCRIPTION	USA ASTM LEVEL 3 (HIGHEST) VIRAL FILTRATION EFFICIENCY : >=99.9% BACTERIAL FILTRATION EFFICIENCY: >=99.9% PARTICLE FILTRATION EFFICIENCY : >= 99.95%% Viral Filtration Efficiency >=99.9% PM 0.1 MICRON FILTRATION FLUID RESISTANCE:>= 160mmHg BREATHABILITY -DELTA P:<5.0mmH20/cm2 FLAMMABILITY : CLASS 1 PRODUCT IS HYPOALLERGENIC,LATEX AND FIBERGLASS-FREE RECOMMENDED FOR SENSITIVE SKIN SOFT & LIGHTWEIGHT
DEVICE INTENDED PURPOSE	This product is intended for general purpose infection control practices. When properly worn, it reduces potential contact by the wearer to blood and bodily fluids
VALIDITY DATE OF REGISTRATION	08/06/2023 - 07/06/2028

LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	3D AIR FLIP MEDICAL FACE MASK	MS1919
2	3D AIR FLIP MEDICAL FACE MASK	MS1314
3	3D AIR FLIP MEDICAL FACE MASK	MS1319