



DEVICE DETAILS

| | |
|----------------------------------|---|
| NAME OF DEVICE | 3D Air Flip Medical Face Mask 4PLY |
| ESTABLISHMENT NAME | JINGGA ANGGUN SDN BHD |
| ROLE OF ESTABLISHMENT | MANUFACTURER |
| REGISTRATION NO | GA2343823-133955 |
| BRAND NAME | MIOS |
| MEDICAL DEVICE CATEGORY | SINGLE-USE DEVICES |
| DEVICE GROUPING TYPE | FAMILY |
| DEVICE DESCRIPTION | USA ASTM LEVEL 3 (HIGHEST) VIRAL FILTRATION EFFICIENCY : >=99.9% BACTERIAL FILTRATION EFFICIENCY: >=99.9% PARTICLE FILTRATION EFFICIENCY : >= 99.95%% Viral Filtration Efficiency >=99.9% PM 0.1 MICRON FILTRATION FLUID RESISTANCE:>= 160mmHg BREATHABILITY -DELTA P:<5.0mmH2O/cm2 FLAMMABILITY : CLASS 1 PRODUCT IS HYPOALLERGENIC,LATEX AND FIBERGLASS-FREE RECOMMENDED FOR SENSITIVE SKIN SOFT & LIGHTWEIGHT |
| DEVICE INTENDED PURPOSE | This product is intended for general purpose infection control practices. When properly worn, it reduces potential contact by the wearer to blood and bodily fluids |
| VALIDITY DATE OF REGISTRATION | 08/06/2023 - 07/06/2028 |

LIST OF DEVICE

MEDICAL DEVICE AUTHORITY

| NO | NAME OF DEVICE | IDENTIFIER |
|----|-------------------------------|------------|
| 1 | 3D AIR FLIP MEDICAL FACE MASK | MS1919 |
| 2 | 3D AIR FLIP MEDICAL FACE MASK | MS1314 |
| 3 | 3D AIR FLIP MEDICAL FACE MASK | MS1319 |