

DEVICE DETAILS		
NAME OF DEVICE	FINGERTIP PULSE OXIMETER	
ESTABLISHMENT NAME	LAWMANN TRADING (SARAWAK) SDN BHD	
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE	
REGISTRATION NO	GC767811121718	
BRAND NAME	BM1000D	
MEDICAL DEVICE CATEGORY	MD 1302 - Monitoring devices of vital physiological parameters	
DEVICE GROUPING TYPE	SINGLE	
FINGERTIP PULSE OXIMETER IS AN IMPORTANT AND COMMON DEVICE TO CHECK OXYGEN SATURATION (SPO2) AND PULSE FIT´ S A SMALL, COMPACT, SIMPLE, RELIABLE AND DURAB PHYSIOLOGICAL MONITORING DEVICE. INCLUDE THE MAINBOA OLED DISPLAY AND DRY BATTERIES. THE PULSE OXIMETER IS REUSE DEVICE AND INTENDED USE FOR SPOT CHECKING OF THE PULSE OXYGEN SATURATION AND PULSE RATE FOR ADULT IN CLINIC ENVIRONMENT. THIS MEDICAL DEVICE CAN BE REUSED FOR CONTINUOUSLY MONITORING.		
DEVICE INTENDED PURPOSE	THE PULSE OXIMETER IS REUSE DEVICE AND INTENDED USE FOR SPOT CHECKING OF THE PULSE OXYGEN SATURATION AND PULSE RATE FOR ADULT IN CLINIC ENVIRONMENT. THIS MEDICAL DEVICE CAN BE REUSED. NOT FOR CONTINUOUSLY MONITORING.	
VALIDITY DATE OF REGISTRATION	26/09/2018 - 25/09/2023	

LIST OF DEVICE			
NO	NAME OF DEVICE	IDENTIFIER	
1	FINGERTIP PULSE OXIMETER	BM1000D	