



## DEVICE DETAILS

<b>NAME OF DEVICE</b>	FINGERTIP PULSE OXIMETER
<b>ESTABLISHMENT NAME</b>	SYMBIOMED SDN. BHD
<b>ROLE OF ESTABLISHMENT</b>	AUTHORIZED REPRESENTATIVE
<b>REGISTRATION NO</b>	GC287471101918
<b>BRAND NAME</b>	ALVITAL
<b>MEDICAL DEVICE CATEGORY</b>	MD 1300 - MONITORING DEVICES
<b>DEVICE GROUPING TYPE</b>	SINGLE
<b>DEVICE DESCRIPTION</b>	FINGER OXIMETRY FOR NON-INVASIVE SPOT CHECKING AND CONTINUOUS MONITORING IN HOME AND PROFESSIONAL CARING ENVIRONMENT, AND TRANSPORT AND MOBILE ENVIRONMENT. IT IS DESIGNED FOR FINGERS.
<b>DEVICE INTENDED PURPOSE</b>	AT101 IS INTENDED FOR MEASURING FUNCTIONAL OXYGEN SATURATION OF ARTERIAL HEMOGLOBIN (SPO2) AND PULSE RATE FOR BOTH ADULTS AND CHILDREN AS NON-INVASIVE SPOT CHECKING AND CONTINUOUS MONITORING IN HOME AND PROFESSIONAL CARING ENVIRONMENT, AND TRANSPORT AND MOBILE ENVIRONMENT. IT IS DESIGNED FOR FINGERS BETWEEN 0.8CM AND 2.3CM (0.3 INCH ~ 0.9 INCH) AND INTEND FOR PATIENTS DURING NO-MOTION CONDITION
<b>VALIDITY DATE OF REGISTRATION</b>	26/09/2018 - 25/09/2023

## LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	FINGERTIP PULSE OXIMETER	AT101