



DEVICE DETAILS

NAME OF DEVICE	FINGERTIP PULSE OXIMETER
ESTABLISHMENT NAME	SYMBIOMED SDN. BHD
ROLE OF ESTABLISHMENT	AUTHORIZED REPRESENTATIVE
REGISTRATION NO	GC287471101918
BRAND NAME	ALVITAL
MEDICAL DEVICE CATEGORY	MD 1300 - MONITORING DEVICES
DEVICE GROUPING TYPE	SINGLE
DEVICE DESCRIPTION	FINGER OXIMETRY FOR NON-INVASIVE SPOT CHECKING AND CONTINUOUS MONITORING IN HOME AND PROFESSIONAL CARING ENVIRONMENT, AND TRANSPORT AND MOBILE ENVIRONMENT. IT IS DESIGNED FOR FINGERS.
DEVICE INTENDED PURPOSE	AT101 IS INTENDED FOR MEASURING FUNCTIONAL OXYGEN SATURATION OF ARTERIAL HEMOGLOBIN (SPO2) AND PULSE RATE FOR BOTH ADULTS AND CHILDREN AS NON-INVASIVE SPOT CHECKING AND CONTINUOUS MONITORING IN HOME AND PROFESSIONAL CARING ENVIRONMENT, AND TRANSPORT AND MOBILE ENVIRONMENT. IT IS DESIGNED FOR FINGERS BETWEEN 0.8CM AND 2.3CM (0.3 INCH ~ 0.9 INCH) AND INTEND FOR PATIENTS DURING NO-MOTION CONDITION
VALIDITY DATE OF REGISTRATION	26/09/2018 - 25/09/2023



LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	FINGERTIP PULSE OXIMETER	AT101