



## DEVICE DETAILS

<b>NAME OF DEVICE</b>	PULSE OXIMETER
<b>ESTABLISHMENT NAME</b>	JITRON HEALTHCARE SDN BHD
<b>ROLE OF ESTABLISHMENT</b>	AUTHORIZED REPRESENTATIVE
<b>REGISTRATION NO</b>	GC8919924-173487
<b>BRAND NAME</b>	JUMPER
<b>MEDICAL DEVICE CATEGORY</b>	MD 1302 - Monitoring devices of vital physiological parameters
<b>DEVICE GROUPING TYPE</b>	FAMILY
<b>DEVICE DESCRIPTION</b>	The fingertip pulse Oximeter features small size, low power consumption, convenient operation, and portability. It is only necessary for a user to put one of his fingers into the fingertip photoelectric sensor for diagnosis, and a display screen will show oxygen saturation. IT IS NOT USED FOR CONTINUOUS MONITORING AND IS NOT USED TOGETHER WITH OTHER MONITORING DEVICES.
<b>DEVICE INTENDED PURPOSE</b>	JPD-500A The Fingertip Pulse Oximeter is non-invasive device intended for spot-checking of functional oxygen saturation of arterial hemoglobin (Spo2) and pulse rate. The portable fingertip device is indicated for adult and pediatric patients in home and hospital environments. JPD-500E The Fingertip Pulse Oximeter is a non-invasive device intended for the spot-check of oxygen saturation of arterial hemoglobin (SpO2) and the pulse rate of adult and pediatric patients in home and hospital environments.
<b>VALIDITY DATE OF REGISTRATION</b>	05/06/2024 - 04/06/2029

## LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	JUMPER FINGERTIP PULSE OXIMETER (500E)	JPD-500E
2	JUMPER PULSE OXIMETER JPD-500A	JPD-500A