



DEVICE DETAILS

NAME OF DEVICE	PULSE OXIMETER
ESTABLISHMENT NAME	TEEPHAM MEDICAL SDN. BHD.
REGISTRATION NO	GB82088101816
BRAND NAME	CA-MI O2 EASY (REF OX 100000)
DEVICE GROUPING TYPE	SINGLE
DEVICE DESCRIPTION	<p>O2-EASY IS A FINGER PULSE OXIMETER TO CHECK AND MONITOR PATIENT BLOOD-OXYGEN SATURATION (SPO2) LEVEL AND PULSE RATE. ITS SMALL, COMPACT, SIMPLE, RELIABLE AND DURABLE DESIGN GREATLY ENHANCES PATIENT CARE. STANDARD ACCESSORIES INCLUDED: 2 X AAA 1.5V ALKALINE BATTERIES (REPLACEABLE), 1 X HANG LACE. THE CA-MI O2 EASY PULSE OXIMETER IS EASY AND PRACTICAL TO USE. IT IS SMALL, LIGHTWEIGHT AND EASY TO CARRY AROUND. THE DEVICE HAS LOW ENERGY CONSUMPTION AND THE TWO AAA BATTERIES HAVE A LIFE OF 30 HOURS OF CONTINUOUS USE. WHEN THE BATTERY CHARGE IS SO LOW THAT IT MAY AFFECT NORMAL FUNCTIONING OF THE PULSE OXIMETER, A WARNING IS SHOWN ON THE DISPLAY. THE DEVICE AUTOMATICALLY TURNS OFF WHEN IT DOES NOT RECEIVE ANY SIGNALS FOR MORE THAN 8 SECONDS. THE CA-MI O2 EASY PULSE OXIMETER BATTERIES MUST BE REPLACED WHEN THE WARNING LIGHT COMES ON. ITS SURFACES MUST BE CLEANED BEFORE USING IT FOR DIAGNOSIS. THE BATTERIES MUST BE REMOVED FROM THE BATTERY COMPARTMENT IF THE DEVICE IS NOT USED FOR A LONG PERIOD OF TIME. IT IS ADVISABLE TO STORE THE DEVICE AT A TEMPERATURE BETWEEN -20°C~55°C(-4°F~131°F) WITH RELATIVE HUMIDITY <93%. THE DEVICE MUST ALWAYS BE STORED IN A DRY ENVIRONMENT. A HUMID ENVIRONMENT MAY DAMAGE THE DEVICE, REDUCING ITS LIFE.</p>
DEVICE INTENDED PURPOSE	TO CHECK AND MONITOR PATIENT BLOOD-OXYGEN SATURATION (SPO2) LEVEL AND PULSE RATE. WIDELY USED IN DIFFERENT MEDICAL FIELDS, SUCH AS HOME-CARE, CLINICS, HOSPITALS, SOCIAL MEDICAL ORGANIZATIONS, FIRST AID, DENTAL OPERATION.
VALIDITY DATE OF REGISTRATION	06/12/2016 - 05/12/2021

LIST OF DEVICE

NO	NAME OF DEVICE	IDENTIFIER
1	PULSE OXIMETER	OX 100000