

DEVICE DETAILS		
NAME OF DEVICE	CHECKME O2 PULSE OXIMETER	
ESTABLISHMENT NAME	BMEC (MALAYSIA) SDN. BHD	
REGISTRATION NO	GB3693921-57628	
BRAND NAME	OXIBAND	
DEVICE GROUPING TYPE	SYSTEM	
DEVICE DESCRIPTION	The CheckMe O2 Pulse Oximeter measures the oxygen saturation and pulse rate. It works by shining two beams into the small blood vessels or capillaries of finger, reflecting the amount of oxygen in the blood and displaying the measurement.	
DEVICE INTENDED PURPOSE	The Oxiband Pulse Oximeter is a wrist pulse oximeter indicated for use in measuring, displaying, storing and transmitting functional oxygen saturation of arterial hemoglobin (SpO2) and pulse rate for adult patients. The Oxiband it is intended for spot-check and/ or continuous data collection, and not continuous monitoring. It can be used in sleep labs, long-term care, hospitals and home use.	
VALIDITY DATE OF REGISTRATION	10/03/2021 - 09/03/2026	

LIST O	F DEVICE Medical Device Autropation	
NO	NAME OF DEVICE MEDICAL DEVICE AUTHORIT	YIDENTIFIER
1	Pulse Oximeter	Oxiband
2	SpO2 Ring Sensor	FP10R
3	USB Charging Cable	540-00240-00