



DEVICE DETAILS

| | |
|-------------------------------|---|
| NAME OF DEVICE | MEDISANA PULSE OXIMETER PM 100 CONNECT |
| ESTABLISHMENT NAME | ZUELLIG PHARMA SDN.BHD |
| REGISTRATION NO | GC9976721-74944 |
| BRAND NAME | MEDISANA |
| DEVICE GROUPING TYPE | SINGLE |
| DEVICE DESCRIPTION | The Pulse Oximeter PM 100 CONNECT is a portable non-invasive device used to measure and record the oxygen saturation of arterial hemoglobin (SpO2) and pulse rate of users at any moment. It is a small and compact device which is applicable in the hospital and home use with simple and convenient operation. However, it is reminded that the Pulse Oximeter PM 100 CONNECT is not suitable for continuous blood oxygen monitoring. The measuring range for the SpO2 is 70% to 99% with 1% resolution while the measuring range for the pulse rate is 30 to 235 beats per minute (bpm) with a resolution of 1 bpm. |
| DEVICE INTENDED PURPOSE | The Pulse Oximeter PM 100 CONNECT is intended for spot-checking of oxygen saturation of arterial hemoglobin (SpO2) and pulse rate users. It is not suitable for continuous monitoring. |
| VALIDITY DATE OF REGISTRATION | 20/09/2021 - 19/09/2026 |

LIST OF DEVICE

MEDICAL DEVICE AUTHORITY

| NO | NAME OF DEVICE | IDENTIFIER |
|----|--|------------|
| 1 | MEDISANA PULSE OXIMETER PM 100 CONNECT | 48506 |