



## DEVICE DETAILS

NAME OF DEVICE	3D Air Flip Face Mask 3PLY
ESTABLISHMENT NAME	JINGGA ANGGUN SDN BHD
REGISTRATION NO	GA2323723-123394
BRAND NAME	MIOS
DEVICE GROUPING TYPE	FAMILY
DEVICE DESCRIPTION	USA ASTM LEVEL 3 (HIGHEST) VIRAL FILTRATION EFFICIENCY : >=99.9% BACTERIAL FILTRATION EFFICIENCY: >=99.9% PARTICLE FILTRATION EFFICIENCY : >= 99.95%% Viral Filtration Efficiency >=99.9% PM 0.1 MICRON FILTRATION FLUID RESISTANCE:>= 160mmHg BREATHABILITY -DELTA P:<5.0mmH2O/cm2 FLAMMABILITY : CLASS 1 PRODUCT IS HYPOALLERGENIC,LATEX AND FIBERGLASS- FREE RECOMMENDED FOR SENSITIVE SKIN SOFT & LIGHTWEIGHT
DEVICE INTENDED PURPOSE	This product is intended for general purpose infection control practices. When properly worn, it reduces potential contact by the wearer to blood and bodily fluids
VALIDITY DATE OF REGISTRATION	10/03/2023 - 09/03/2028

## LIST OF DEVICE



NO	NAME OF DEVICE	MEDICAL DEVICE AUTHORITY	IDENTIFIER
1	3D AIR FLIP MEDICAL FACE MASK		MS1313
2	3D AIR FLIP MEDICAL FACE MASK		MS1414
3	3D AIR FLIP MEDICAL FACE MASK		MS9131